

Nitrile disposable Gloves | POWDER-FREE



MEDCARE
Medical protection

QUALITY STANDARDS

- Conforms to EN455, EN 420, EN374-1, EN374-2:3019 and EN ISO 21420:2020
- Manufactured under QSR (GMP), ISO 9001:2015 Quality Management System

GLOVE SIZES

- Small, Medium, Large, Extra-Large
- Size of gloves shall be marked in the check box on the shipping carton with black ink

PRODUCT SPECIFICATIONS

Type	Powdered & Powder-Free, Non-sterile
Material	100% Nitrile Latex-Free
Colour	Blue
Design & Features	Powdered-Free: Polymer coated or online single chlorinated, offline double chlorinated, ambidextrous, finger textured or palm textured surface, beaded cuff
Storage	The gloves shall maintain their properties when stored in a dry condition. Avoid direct sunlight
Shelf-Life	3 years from the date of manufacturing



PRODUCT SPECIFICATIONS

Standard			
EN 455			

Length (mm)	Min 230,	Min 220 (S)	Min 240
	Min 240,	Min 230 (M, L, XL)	
	300 ± 10		

Palm Width (mm)			
• S	84 ± 3	80 ± 10	80 ± 10
• M	94 ± 3	95 ± 10	95 ± 10
• L	105 ± 3	110 ± 10	110 ± 10
• XL	113 ± 3	120 ± 10	2:110

Thickness: Single Wall (mm)			
• Finger	Min 0.08 -0.12	Min 0.08	N/A
• Palm	Min 0.05 -0.12	Min 0.05	N/A

PHYSICAL PROPERTIES


Property	EN 455	
----------	--------	--

Tensile Strength (MPa)		
• Before Aging	14	N/A
• After Aging	14	N/A

Elongation at Break (%)		
• Before Aging	Min 500	N/A
• After Aging	Min 400	N/A

Median Force at Break (N)		
• Before Aging	N/A	Min 6
• After Aging	N/A	Min 6

MEDCARE
Medical protection





2020

CERTIFICATE OF REGISTRATION

This certifies that:

DONGGUAN GRINVALD TECHNOLOGY CO. LTD
401, Building #3, No 4 Of Guangming New Village 2 Road
Dongcheng
Dongguan City Guangdong, CN 523000

is registered with the U.S. Food and Drug Administration for FY 2020 pursuant to Title 21, 807 et seq. of the United States Code of Federal Regulations:

Establishment Owner/Operator Number:	10071856
DUNS No.:	55-453-3470
Device Classification Name:	POLYMER PATIENT EXAMINATION GLOVE
Product Code:	LZA
Regulation Number:	880.6250
Official Correspondent and U.S. Agent:	Registrar Corp 144 Research Drive, Hampton, Virginia, 23666, USA Telephone: +1-757-224-0177 • Fax: +1-757-224-0179

Registrar Corp will confirm that such registration remains effective upon request and presentation of this certificate until the end of the year stated above, unless said registration is terminated after issuance of this certificate. Registrar Corp makes no other representations or warranties, nor does this certificate make any representations or warranties to any person or entity other than the named certificate holder, for whose sole benefit it is issued. This certificate does not denote endorsement or approval of the certificate-holder's device or establishment by the U.S. Food and Drug Administration. Registrar Corp assumes no liability to any person or entity in connection with the foregoing.

Pursuant to 21 CFR 807.39, "Registration of a device establishment or assignment of a registration number does not in any way denote approval of the establishment or its products. Any representation that creates an impression of official approval because of registration or possession of a registration number is misleading and constitutes misbranding."

The U.S. Food and Drug Administration does not issue a certificate of registration, nor does the U.S. Food and Drug Administration recognize a certificate of registration. Registrar Corp is not affiliated with the U.S. Food and Drug Administration.

Registrar Corp
144 Research Drive, Hampton, Virginia, 23666, USA
Telephone: +1-757-224-0177 • Fax: +1-757-224-0179
info@registrarcorp.com • www.registrarcorp.com

David Lennarz
Executive Director
Registrar Corp

Dated: July 13, 2020



TÜVRheinland®
Precisely Right.

Products

Report No.:	242122000-01a	Page 1 of 3
Client:	Dongguan Grinvald Technology Co., Ltd.	
Contact Information:	401 Building A 4th Floor, Dongguan Guangdong China 523000	
Identification/ Model No(s):	MEDCARE NITRILE EXAMINATION GLOVES	
Sample Receiving date:	2020-08-05	
Sample Resubmitted date:	2020-08-20	
Testing Period:	2020-08-05 to 2020-08-27	
Delivery condition:	Apparent good, Samples tested as received	

Test Specification:	Test result:
1. EN 455-1: 2000: Requirements for freedom from holes	PASS

Other Information provided by client:

Grade: Examination Gloves Powder Free
Manufacture: Dongguan Grinvald Technology Co., Ltd.
Country of Origin: China

The report 242122000-01a supersedes report 242122000-01 (Revised Identification/Model No.(S))

For and on behalf of
TÜV Rheinland Thailand Ltd.



Wilawan Sripthrom / Manager

2020-09-01

Date
Name/Position
Test result is drawn according to the kind and extent of tests performed. This test report relates to the above mentioned Test sample. Without permission of the test center this test report is not permitted to be duplicated in extracts. This test report does not entitle to carry any safety mark on this or similar products.

TÜV Rheinland Thailand Ltd. - Global Technology Assessment Center Bangkok (GTAC BKK) Ladkrabang Industrial Estate 123/1, Soi Chalokkung 31, Lamplataw, Ladkrabang, Bangkok 10520 Thailand
Tel.: +66 (0) 2326-1333 Fax: +66 (0) 2326-1334-5 Email: info@tha.tuv.com Web: www.tuv.com

Test Report No.: 242122000-01a

Page 2 of 3

Sampling Information:

Inspection Method: No inspection
 Inspection level: N/A
 AQL: N/A
 Sample size: N/A

Material list:

Material No.	Material	Color	Location
M001	Nitrile Gloves	Blue	Refer to photo

Freedom from holes

Test method: With reference to EN 455-1: 2000

Test result:

Material No.	Gloves Size	Tested samples	No. of samples for Non-compliance	Conclusion
M001	M	200 pcs.	1	Pass

Remark:

- All samples were selected and supplied by the client.
- The batch size of the gloves supplied was not stated by the client. In accordance with BS EN 455-1, a batch size between 35,001 to 150,000 was chosen, and therefore 50 gloves per stage were tested for perforations using General Inspection Level I at an AQL of 1.5%. with reference to table, the result can be judged as above AQL 0.65.

Stage No.	Cumulative no. tested	Accept	Reject
First	50	0	4
Second	100	1	6
Third	150	3	8
Fourth	200	5	9
Fifth	250	9	19

Test Report No.: 242122000-01a

Page 3 of 3

Sample photo

-END-

Sample Photo


The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

II. The content of this part is extracted from the test report number GZHL20070373310T.
1. SGS Ref No.: CAN20-117986

Test Requested	Result
Entry 63 of Commission Regulation (EU) 2015/628 amending Annex XVII of REACH Regulation (EC) No 1907/2006 - Lead and its compounds	PASS
Entry 23 of Commission Regulation (EU) 2016/217 amending Annex XVII of REACH Regulation (EC) No 1907/2006 - Cadmium and its compounds	PASS
Entry 50 of Commission Regulation (EU) 2015/326 amending Annex XVII of REACH Regulation (EC) No 1907/2006 - Polycyclic Aromatic Hydrocarbons (PAHs)	PASS
Entry 51 of Commission Regulation (EU) 2018/2005 amending Annex XVII of REACH Regulation (EC) No 1907/2006 - Phthalates	PASS
European Regulation POP's (EU) 2019/1021 - Alkanes C10-C13, chloro (short-chain chlorinated paraffins) (SCCPs)	PASS



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/sgs-conditions-service.pdf>, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/sgs-conditions-service-electronic-format-documents.pdf>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced or used in any form without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only. Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8387 1843, or email: CN_Client@sgs.com

19# Kechu Road, Science Park, Guangzhou Economic & Technology Development District, Guangzhou, China 510663 | (86-20) 82155555 | (86-20) 82075191 | www.sgs.com.cn
 中国·广州·经济技术开发区科学城科珠路190号 邮编: 510663 | (86-20) 82155555 | (86-20) 82075191 | sgs.china@sgs.com

Member of the SGS Group (SGS SA)

Test Results :
Test Part Description :

SGS Sample ID	Description
CAN20-117986.001	Blue soft plastic(glove)

Remarks :

- (1) 1 mg/kg = 0.0001%
- (2) MDL = Method Detection Limit
- (3) ND = Not Detected (< MDL)
- (4) "-" = Not Regulated

Entry 63 of Commission Regulation (EU) 2015/628 amending Annex XVII of REACH Regulation (EC) No 1907/2006 - Lead and its compounds

Test Method : SGS In-house method (GZTC CHEM-TOP-004-01, with reference to EPA 3052:1996), analysis was performed by ICP-OES.

Test Item(s)	Limit	Unit	MDL	001
Lead (Pb)	500	mg/kg	2	ND

Entry 23 of Regulation (EU) 2016/217 amending Annex XVII of REACH Regulation (EC) No 1907/2006- Cadmium and its compounds

Test Method : SGS In-house method (GZTC CHEM-TOP-004-01, with reference to US EPA Method 3052:1996), analysis was performed by ICP-OES.

Test Item(s)	CAS NO.	Limit	Unit	MDL	001
Cadmium (Cd)	7440-43-9	0.01	%(w/w)	0.0005	ND

Entry 50 of Regulation(EU) 2015/326 amending Annex XVII of REACH Regulation (EC) No 1907/2006- Polycyclic Aromatic Hydrocarbons (PAHs)

Test Method : With reference to AIPS GS 2019:01 PAK, analysis was performed by GC-MS.

Test Item(s)	CAS NO.	Limit	Unit	MDL	001
Benzo(a)anthracene(BaA)	56-55-3	1.0	mg/kg	0.1	ND
Chrysene(CHR)	218-01-9	1.0	mg/kg	0.1	ND



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/sgs-conditions-service.pdf>, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/sgs-conditions-service-electronic-format-documents.pdf>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced or used in any form without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only. Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8387 1843, or email: CN_Client@sgs.com

19# Kechu Road, Science Park, Guangzhou Economic & Technology Development District, Guangzhou, China 510663 | (86-20) 82155555 | (86-20) 82075191 | www.sgs.com.cn
 中国·广州·经济技术开发区科学城科珠路190号 邮编: 510663 | (86-20) 82155555 | (86-20) 82075191 | sgs.china@sgs.com

Member of the SGS Group (SGS SA)

WORK REQUESTED

Samples described as Nitrile examination gloves, powder free, colour blue, referenced MD0120. Size S (6), M (7), L (8) were received by SATRA on 22 July 2020 for testing in accordance with EN ISO 21420: 2020 and EN 374-2: 2014.

SAMPLE SUBMITTED



Samples described as Nitrile examination gloves, powder free, colour blue, referenced MD0120. Size S (6), M (7), L (8).

TESTING REQUESTED

EN ISO 21420: 2020 Clause 5.1 – Sizing and measurement of gloves
 EN ISO 21420: 2020 Clause 5.2 – Dexterity
 EN ISO 21420: 2020 Clause 4.2 – Innocuousness of protective gloves
 EN 374-2: 2014 Clause 7.2 – Air leak
 EN 374-2: 2014 Clause 7.3 – Water leak

CONCLUSION

The samples described as Nitrile examination gloves, powder free, colour blue, referenced MD0120. Size S (6), M (7), L (8) were found to achieve the following results:

EN ISO 21420: 2020 Clause 5.1 – See below table
 EN ISO 21420: 2020 Clause 5.2 – Level 5
 EN ISO 21420: 2020 Clause 4.2* – Pass PAHs, DMFa and pH value
 EN 374-2: 2014 Clause 7.2* – Pass
 EN 374-2: 2014 Clause 7.3 – Pass

All tests marked * in this technical report were subcontracted to test facilities accredited to ISO/IEC 17025: 2017 by CNAS.

Detailed results are included on the following page(s)

Testing

Testing was carried out in accordance with EN 374-2: 2014

Samples for testing were conditioned for at least 24 hours in a conditioned environment maintained at (23±2)*C and (50±5)% relative humidity.

Requirements

Requirements for EN 374-2: 2014

Clause 7.2 Air leak	No leak to be detected
Clause 7.3 Water leak	No leak to be detected

Test Results

EN 374-2: 2014 Test Results

Clause / Test	Test Results	UoM	Result
7.2 Air leak test	Total air pressure used	3.0 kPa	NA
	Sample size	Leaks	
	6	No leaks detected	
	7	No leaks detected	
	8	No leaks detected	
7.3 Water leak test	Sample size	Leaks	NA
	6	No leaks detected	
	7	No leaks detected	
	8	No leaks detected	
	8	No leaks detected	

*** End of Report ***



March 9, 2016

Zhonghong Pulin Medical Products Co., Ltd.
c/o Mr. Chu Xiaon
Room 1606 Bldg. 1, Jianxiang Yuan No. 209
Bei Si Huan Zhong Road, Haidian District
Beijing 100083
CHINA

Re: K152712

Trade/Device Name: Nitrile Powder Free Patient Examination Gloves, Blue Color
Regulation Number: 21 CFR 880.6250
Regulation Name: Patient Examination Glove
Regulatory Class: Class I
Product Code: LZA
Dated: January 28, 2016
Received: February 1, 2016

Dear Mr. Xiaon:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the [Federal Register](#).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in

Page 2 - Mr. Xiaon

the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Tejashri Purohit-Sheth, M.D.

Tejashri Purohit-Sheth, M.D.
Clinical Deputy Director
DAGRID/ODE/CDRH FOR

Erin I. Keith, M.S.
Director
Division of Anesthesiology,
General Hospital, Respiratory,
Infection Control, and Dental Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES Food and Drug Administration Indications for Use	Form Approved: OMB No. 0910-0120 Expiration Date: January 31, 2017 See PRA Statement below.
---	---

510(k) Number (if known)

K152712

Device Name

Nitrile Powder Free Patient Examination Gloves, Blue Color

Indications for Use (Describe)

Nitrile Powder Free Patient Examination Gloves, Blue Color is a non-sterile disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D) Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

	powder	2mg of residual powder	
Materials used to fabricate the devices	Nitrile	Nitrile	Substantially equivalent
Dusting or Donning Powder: name	PU	Polyurethane	Substantially equivalent
Compare performance data supporting substantial equivalence	Meets ASTM D5151-06 (Reapproved 2011) ASTM D6319-10 ASTM D6124-06 (Reapproved 2011)	Meets ASTM D5151-06 (Reapproved 2011) ASTM D6319-10 ASTM D6124-06 (Reapproved 2011)	Substantially equivalent
Single Patient Use	Single Patient Use	Single Patient Use	Substantially equivalent
Biocompatibility	SKIN IRRITATION DERMAL and SENSITIZATION STUDIES Meets ISO 10993-10:2002/Amd.1:2006 Under the conditions of the study, not an irritant and under conditions of the study, not a sensitizer.	SKIN IRRITATION DERMAL and SENSITIZATION STUDIES Meets ISO 10993-10: Third Edition 2010-08-01. Under the conditions of the study, not an irritant and under conditions of the study, not a sensitizer.	Substantially equivalent
Labeling for the legally marketed device to which substantial equivalence is claimed.	-Powder Free -Patient Examination Glove -Single Use Only - Manufactured For: - Lot -Blue color - Non sterile	-Powder Free -Patient Examination Glove -Single Use Only - Manufactured For: - Lot -Blue color - Non sterile	Substantially equivalent

10.0 Substantial Equivalence Comparison:

It can be concluded that the Nitrile Powder Free Patient Examination Gloves, Blue Color meet the ASTM standard or equivalent standard and FDA requirements for waterleak test on pinhole AQL, meet labeling claims.

It can be concluded that the Nitrile Powder Free Patient Examination Gloves, Blue Color is safe, as effective, and performs as well as the predicate device, Nitrile Powder Free Patient Examination Gloves, Blue Color, Tangshan Zhonghong Pulin Plastic Co., Ltd. K120970.

